

Issued by NMI Certin B.V.

In accordance with WELMEC 8.8 2017, WELMEC 2.4 2021, OIML R 60 (2021), EN 45501:2015.

Producer Anyload Youngzon Transducer (Hangzhou) Co., Ltd
518, 18th Street, Qiantang New Area
Hangzhou
China

Measuring instrument **A bending beam load cell (planar beam)**, with strain gauges, tested as a part of a weighing instrument.

Registered trade name : ANYLOAD
Designation : 202UA

Further properties are described in the annexes:

- Description TC12265 revision 0;
- Documentation folder TC12265-1.

An overview of performed tests is given in the annex:

- Description TC12265 revision 0.

Initially issued 28 December 2023

Issuing Authority

NMI Certin B.V.
28 December 2023

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the producer shall indemnify third-party liability.

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.

1 General information about the load cell

All properties of the load cell, whether mentioned or not, shall not be in conflict with the standards mentioned in this certificate.

This certificate is the positive result of the applied voluntary, modular approach, for a component of a measuring instrument, as described in WELMEC 8.8. The complete measuring instrument must be covered by relevant metrological certification that is valid in the country where the instrument is put into use.

1.1 Essential parts

Number	Pages	Description	Remark
12265/0-01	1	Outline drawing	Mechanical
12265/0-02	1	Wiring diagram	Electrical

Cable:

- If the load cell is provided with a 4-wire system:
 - The cable length is mentioned in the accompanying load cell document / on the label;
 - The cable length shall not be modified.
- If the load cell is provided with a 6-wire system (=“Remote-sensing”):
 - The cable length is not limited.

The cable is shielded; the shield is not connected to the load cell.

1.2 Essential characteristics

Characterization of load cell capabilities	Analog-passive load cell
Maximum capacity (E_{max})	75 kg up to and including 375 kg
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	1,0 mV/V \pm 20%
Maximum number of load cell intervals (n) ⁽¹⁾	3000
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{max} / V_{min}$	28000
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{max} / (2 * DR)$	6000
Input impedance	1190 Ω \pm 50 Ω
Temperature range	-10 $^{\circ}$ C / + 40 $^{\circ}$ C
Fraction p_{LC}	0,7
Humidity Class	CH

Safe overload	150 % of E_{max}
Output impedance	$1000 \Omega \pm 10 \Omega$
Recommended excitation	5 - 10 V DC
Excitation maximum	15 V DC
Transducer material	Aluminium
Atmospheric protection	Silicone sealed

Remarks:

1. The characteristics for n_{max} , Y and Z can be reduced separately.

1.3 Essential shapes

Number	Pages	Description	Remark
12265/0-01	1	Outline drawing	Mechanical

The descriptive markings plate is secured against removal by sealing or will be destroyed when removed and contains at least the information and markings as described in OIML R 60 (2021) and:

- This certificate number TC12265 (in the countries where it is mandatory);
- Producers name or mark.

2 Seals

The connecting cable of the load cell or the junction box is provided with possibility to seal.

3 Conditions for conformity assessment

Each load cell produced is provided with an accompanying document with information about its characteristics.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN45501:2015 clause F.4, at the time of putting into use.

Other parties may use this certificate without the written permission of the producer.

4 Reports

An overview of performed tests is given in the evaluation report ER12265 revision 0.