

Issued by NMI Certin B.V.

In accordance with WELMEC 8.8 2017, WELMEC 2.4 2021, OIML R 60 (2021), EN 45501:2015.

Producer Anyload Youngzon Transducer (Hangzhou) Co.Ltd  
518, 18th Street, Qiangtang New Area  
Hangzhou  
China

Measuring instrument A **shear beam load cell**, with strain gauges, tested as a part of a weighing instrument.

Registered trade name : ANYLOAD  
Designation : 563Wxxx

Further properties are described in the annexes:

- Description TC12317 revision 1;
- Documentation folder TC12317-1.

An overview of performed tests is given in the annex:

- Description TC12317 revision 1.

Remark This revision replaces the earlier version, except for its documentation folder.

Issuing Authority **NMI Certin B.V.**  
19 January 2023

Certification Board

## 1 General information about the load cell

All properties of the load cell, whether mentioned or not, shall not be in conflict with the standards mentioned in this certificate.

This certificate is the positive result of the applied voluntary, modular approach, for a component of a measuring instrument, as described in WELMEC 8.8. The complete measuring system must be covered by an EC type-approval certificate, an EU-type examination certificate, or an approval that is valid in the country where the load cell is taken into service.

### 1.1 Essential parts

Number	Pages	Description	Remark
12317/0-01	1	Outline drawing	Mechanical
12317/0-02	1	Outline drawing measuring element	-
12317/0-03	1	Electrical drawing	-

Cable:

- If the load cell is provided with a 4-wire system:
  - The cable length is mentioned in the accompanying load cell document;
  - The cable length shall not be modified.
- If the load cell is provided with a 6-wire system (=“Remote-sensing”):
  - The cable length is not limited.

The cable is shielded; the shield is not connected to the load cell.

### 1.2 Essential characteristics

Characterization of load cell capabilities	Analog-passive load cell
Maximum capacity ( $E_{max}$ )	9000 kg up to and including 45000 kg
Minimum dead load	0 kg
Accuracy Class	C
Rated Output	$3,0 \pm 0,3$ mV/V
Maximum number of load cell intervals (n) <sup>(1)</sup>	6000
Ratio of minimum LC Verification interval <sup>(1)</sup> $Y = E_{max} / V_{min}$	25000
Ratio of minimum dead load output return <sup>(1)</sup> $Z = E_{max} / (2 * DR)$	6000
Input impedance	$400 \Omega \pm 50 \Omega$

Temperature range	-10 °C / + 40 °C
Fraction $p_{LC}$	0,7
Humidity Class	CH
Safe overload	150 % of $E_{max}$
Output impedance	350 $\Omega \pm 5 \Omega$
Recommended excitation	10 V AC / DC
Excitation maximum	15 V AC / DC
Transducer material	Stainless steel
Atmospheric protection	Silicone sealed

Remark:

1. The characteristics for  $n_{max}$ , Y and Z can be reduced separately.

### 1.3 Essential shapes

Number	Pages	Description	Remark
12317/0-01	1	Outline drawing	Mechanical

The descriptive markings plate is secured against removal by sealing or will be destroyed when removed and contains at least the information and markings as described in OIML R 60 (2021) and:

- This certificate number TC12317 (in the countries where it is mandatory);
- Producers name or mark.

## 2 Seals

The connecting cable of the load cell or the junction box is provided with possibility to seal.

## 3 Conditions for conformity assessment

Each load cell produced is provided with an accompanying document with information about its characteristics.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN45501:2015 clause F.4, at the time of putting into use.

Other parties may use this certificate without the written permission of the producer.

## 4 Reports

An overview of performed tests is given in the evaluation report ER12317 revision 1.